

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Visual Chart System

Name and address of the applicant

View-M Technology Co., Ltd.
4F, 25, Heungan-daero, Gunpo-si, Gyeonggi-do, 15809
Korea, Republic of

Name and address of the manufacturer

View-M Technology Co., Ltd.
4F, 25, Heungan-daero, Gunpo-si, Gyeonggi-do, 15809
Korea, Republic of

Name and address of the factory

Note: When more than one factory, please report on page 2

View-M Technology Co., Ltd.
4F, 25, Heungan-daero, Gunpo-si, Gyeonggi-do, 15809
Korea, Republic of

☐ Additional information on page 2

Ratings and principal characteristics

Input: 100-240 V~, 50/60 Hz, 1.5 A~0.7 A
Output: 12 Vdc, 5.0 A

Trademark / Brand (if any)



View-M

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

VLC-2400P / VLC-2400, CSPOLA600, CSPOLA600C, CSPOLA600F

Additional information (if necessary may also be reported on page 2)

Additionally evaluated in accordance with, IEC 62366-1:2015 + A1:2020.

☐ Additional information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1-6:2010/AMD1:2013, IEC 60601-1-6:2010/AMD2:2020, IEC
60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020

National differences:

CA, US

As shown in the Test Report Ref. No. which forms part of this Certificate

23-074085-02-2

This CB Test Certificate is issued by the National Certification Body

Korea Testing Laboratory (KTL)
87, Digital-ro 26-gil, Guro-gu
Seoul, 08389
Korea, Republic of



Date: 2024-06-11

Signature: Jong Koo Park